



INDIAN INSTITUTE OF TECHNOLOGY BOMBAY
MATERIALS MANAGEMENT DIVISION
Powai, Mumbai 400076.

Ref. PR No. 1000058088

RFx. No. 6100003144

Technical Description: ANALYTICAL ULTRACENTRIFUGE ACCESSORIES (Parts of the existing equipment)

Sr. No	Item Description and Detailed Technical Specification	Technical Compliance (Yes / No)	Additional Information (if any)
1	Centrepiece, Double Sector, Aluminium, 12 mm Part No:334623 QTY : 05		
2	Sapphire Window Replacement – Recommended for Rayleigh Interference Optics System; Not Recommended for Absorbance Work Below 235 nm Part No: 337177 QTY : 07		