INDIAN INSTITUTE OF TECHNOLOGY BOMBAY

MATERIALS MANAGEMENT DIVISION

Powai, Mumbai - 400076

Corrigendum – I

Date: 19.11.2026

PR: 1000058812

RFx. 6100003218

Multi-Point BET Surface Area & Basic Requirements

Sr. No.	Online Rfx Clause	Previous Clause	Changed Clause
1.	Technical Specification	Point 15. Dewar flask Capacity: Cryogenics Dewar Flask with at-least 2L Capacity with Liquid N2 Holding time of 40 hours or more	Point 15. Dewar flask Capacity: Cryogenics Dewar Flask with at-least 2L Capacity with Liquid N2 Holding time of 30 hours or more

Asst. Registrar (MMD)

19/11/2026