



INDIAN INSTITUTE OF TECHNOLOGY BOMBAY

MATERIALS MANAGEMENT DIVISION

Powai, Mumbai - 400076

PR No. 1000038236

Rfx No. 6100001683

Technical Specifications of Add-on Components for Existing NMR 600 MHz Spectrometer

Sr.no	Content	Quantity	Compliance (Yes/No)
1	3.2mm Zirconia Solid State Rotors for existing JEOL 600 MHz ECZR NMR Spectrometer, capable of performing Solid State NMR experiments at MAS ≥ 20 KHz spinning and should be compatible to the existing JEOL 600 MHz ECZR series	5	
2	1 mm Zirconia Solid State Rotors for existing JEOL 600 MHz ECZR NMR Spectrometer, capable of performing Solid State NMR experiments at MAS ≥ 80 KHz spinning and should be compatible to the existing JEOL 600 MHz ECZR series.	10	