



Dated : 27.09.2021

Corrigendum – II

For (PR No. 1000017939) Rfx No. 6100000856

High Purity Nitrogen Gas Generator

Sr. No.	Online Rfx Clause	Previous Clause	Changed Clause
1	Technical specification – point no. 9	Noise level :< 55 dB	The generator should be compact, stackable with an inbuilt air compressor which would have noise less than < 55db.
2	Technical specification – Hydrogen H2 - point no. 1	110 cc/min of hydrogen	Hydrogen capacity to be 100cc or more
3.	Technical specification – Hydrogen H2 - point no. 2	100% titanium PEM cell technology. Purity > 99.9996%	100% titanium/platinum PEM cell technology for Hydrogen with purity > 99.9995%.
4.	Additional technical specification point	<ul style="list-style-type: none"> The nitrogen generator should have HC content <0.05PPM. 	

<p>Digital Signature Deputy Registrar Material Management (misc.drmm.iitb) 28-Sep-21 11:18:24 AM</p> <p>Additional Registrar</p>
